

Boiling Distribution of Petroleum Distillates by Gas Chromatography

ASTM D6352

Configuration of an Agilent Technologies 7890A Series Gas Chromatograph with a flame ionization detector (FID). Method is applicable to petroleum distillates fractions having an initial boiling point greater than 174°C and a final boiling point of 720° C (at 760 torr).