

Boiling Point Distribution of Gasoline and Gasoline Fractions by Gas Chromatography

ASTM D3710

Configuration of an Agilent Technologies 7890N Series Gas Chromatograph with a flame ionization detector (FID). Method is applicable to petroleum products and fractions with a final boiling point of 500°F (260°C) or lower.