

Detailed Hydrocarbon Analysis (DHA) of Petroleum Fractions by Gas Chromatography

ASTM D6730

Configuration of an Agilent Technologies 7890N Series Gas Chromatograph with a flame ionization detector (FID) to perform detailed hydrocarbon analysis.

Performs either of the following methods

- Standard DHA (D6730) – 175 Minute Analysis Time
- Fast DHA – 80-85 Minute Analysis Time *Customer to specify method at time of order.