

Determination of Ethanol Content of Denatured Fuel Ethanol

ASTM D5501-94

This method analyzes 93 to 97 mass % ethanol from 0.1 to 0.6 mass % methanol. This method cannot be used for compounds that have poor or no response in the FID, such as water.