

## GC/MS/PDHID Impurities in Polymer-grade Ethylene and Polymer-grade Propylene

---

Agilent Technologies 7890N Series gas chromatograph configured for use with the Agilent 5973 mass spectroscopy and pulsed discharge helium ionization detectors (MSD/PDHID) for the impurities in polymer-grade ethylene and polymer-grade propylene.

### Determined Components:

• Ar sine	0.05-100 ppm
• Phosphine	0.05-100 ppm
• H <sub>2</sub> S	0.05-100 ppm
• COS	0.01-100 ppm
• Methyl mercaptan	0.05-100 ppm
• Ethyl mercaptan	0.05-100 ppm
• Methanol	0.05-100 ppm
• Ethanol	0.05-100 ppm
• i-Propanol	0.05-100 ppm
• n-Propanol	0.05-100 ppm
• MTBE	0.05-100 ppm
• t-Butanol	0.05-100 ppm
• sec-Butanol	0.05-100 ppm
• n-Butanol	0.05-100 ppm
• Oxygen	1.00-100 ppm
• Nitrogen	1.00-100 ppm
• Hydrogen	1.00-100 ppm