

# Impurities in Chlorine

---

Configuration of an Agilent Technologies 7890N Series Gas Chromatograph with dual thermal conductivity detectors (TCD/TCD) for the Impurities in Chlorine.

## **Determines Components**

- Hydrogen
- Carbon dioxide
- Oxygen/Argon composite
- Nitrogen
- Methane
- Carbon monoxide

Oxygen/Argon composite Lower quantifiable limit is 400 ppm except for carbon monoxide, which is 800 ppm.