

# Standard Boiling Point Distribution of Samples with Residues Such as Crude Oils and Atmospheric and Vacuum Residues by High Temperature Gas Chromatography

---

## ASTM D7169

Configuration of an Agilent Technologies 7890A Series Gas Chromatograph with a flame ionization detector (FID). Method is used to determine the boiling point distribution through a temperature of 720°C. Temperature corresponds to the elution of n-C100. Material boiling above 720°C is reported as residue.